

Unit: micron

Peak %T = 90.2% at 4.342

Half power points: 45.1%T @ 4.278 45.1%T @ 4.434

Half band width = 0.155 % Band width = 3.574%


Center wavelength = 4.356

Test conditions: temp=amb. angle=0 PE983G

Instrument operating parameters: res=3.0cm-1 apert=.8"

Attenuation=0.1%T Below-BP-8.0um

Date: 05/22/91



FILTER SPECIFICATION

- 112 Narrow Band CO/2 Analysis

silverlight^S

Worldwide Product Distribution and Free Technical Support:
Silverlight Ltd., Technoparkstr. 2, CH-8406 Winterthur, Switzerland
info@silverlight.ch www.silverlight.ch